

Original Article

# Effect of Different Surface Treatments on Repair Micro-shear Bond Strength of Silica- and Zirconia-filled Composite Resins

Mohammad Joulaei<sup>1</sup> • Mahmoud Bahari<sup>2,3\*</sup> • Anahid Ahmadi<sup>3</sup> • Siavash Savadi Oskoei<sup>2,4</sup>

<sup>1</sup> Assistant Professor, Department of Operative Dentistry, Faculty of Dentistry, Zanjan University of Medical Sciences, Zanjan, Iran

<sup>2</sup> Dental and Periodontal Research Center, Tabriz Medical Sciences University, Tabriz, Iran

<sup>3</sup> Assistant Professor, Department of Operative Dentistry, Faculty of Dentistry, Tabriz University of Medical Sciences, Tabriz, Iran

<sup>3</sup> Department of Operative Dentistry, Faculty of Dentistry, Zanjan University of Medical Sciences, Zanjan, Iran

<sup>4</sup> Associate Professor, Department of Operative Dentistry, Faculty of Dentistry, Tabriz University of Medical Sciences, Tabriz, Iran

\*Corresponding Author; E-mail: bahari.dds@gmail.com

Received: 8 November 2011; Accepted: 30 October 2012

J Dent Res Dent Clin Dent Prospect 2012; 6(4) :131-137 | doi: 10.5681/joddd.2012.027

This article is available from: <http://dentistry.tbzmed.ac.ir/joddd>

© 2012 The Authors; Tabriz University of Medical Sciences

This is an Open Access article distributed under the terms of the Creative Commons Attribution License (<http://creativecommons.org/licenses/by/3.0>), which permits unrestricted use, distribution, and reproduction in any medium, provided the original work is properly cited.

---

## Abstract

**Background and aims.** Effect of surface treatments on repair bond strength of aged composite resins might be different due to their dissimilar fillers. The aim was to evaluate the effect of different surface treatments on repair micro-shear bond strength ( $\mu$ SBS) of silica- (Spectrum TPH) and zirconia-filled (Filtek Z250) composite resins.

**Materials and methods.** Twenty-seven composite resin blocks were made from each type of composite resin: Z250 and Spectrum TPH. After aging, blocks of each type were randomly divided into three groups according to surface treatments: alloy primer, silane, and only surface roughening. Subsequently, each group was further subdivided into 3 subgroups based on the adhesive system used: Single Bond, Clearfil SE Bond, and Margin Bond. Four composite resin columns were added on each block. After thermocycling,  $\mu$ SBS test were done at cross head speed of 0.5 mm/min. Data was analysed using multifactor ANOVA, one-way ANOVA and a post-hoc Bonferroni tests ( $\alpha = 0.05$ ).

**Results.** Analysis of data showed that the effect of composite resin type was not significant ( $p > 0.05$ ), but the effects of the type of surface treatment ( $p = 0.01$ ) and the type of adhesive system ( $p = 0.01$ ) were significant on repair  $\mu$ SBS. In addition, the cumulative effect of the composite type–surface treatment and the composite type with the type of adhesive system were not statistically significant ( $p > 0.05$ ). However, the cumulative effects of the adhesive system–surface treatment ( $p = 0.03$ ) and the composite type–the adhesive system–surface treatments ( $p = 0.002$ ) were significant.

**Conclusion.** Although repair  $\mu$ SBS values of both silica- and zirconia-filled composite resins were similar, use of different combinations of surface treatments and adhesive systems affected their repair  $\mu$ SBS differently.

**Key words:** Adhesive system, composite resin, etch-and-rinse, filler, micro-shear bond strength, surface treatment, self-etch.

---

## Introduction

One of the most frequent requirements of today's dentistry is bonding of new composite resin to an aged one. It has been estimated that half of a general practitioner's time is spent on replacement dentistry, with the consequent high cost in time and expense.<sup>1</sup> Studies have demonstrated an ideal bonding of two composite resin layers in the presence of an oxygen-inhibited layer.<sup>2,3</sup> However, there is controversy over the effect of oxygen-inhibited layer on composite-composite bond strength.<sup>4,5</sup> In cases with less than 24 hours' time lapse after composite curing, there are enough free and unreacted monomers with reactive vinyl groups ( $C=C$ ) in the composite to copolymerize with vinyl groups of new composite resin to form a strong bond. However, after a long period from composite resin curing, these free unreacted vinyl groups tend to exit the composite, which might result in weak composite-composite repair bond strength.<sup>6</sup>

According to Brosh et al<sup>7</sup> the bonding of old and new composite resin in a repair case might occur by three distinct mechanisms: chemical bonding with the organic matrix, chemical bonding with exposed filler particles, and through micromechanical retention to the treated surface. As mentioned previously a strong chemical bond to resin matrix of an aged composite is questionable and in order to improve repair bond strength, different surface treatments have been used: silica coating, phosphoric acid etching followed by an adhesive,<sup>8</sup> hydrofluoric acid etching, abrasion, and sandblasting.<sup>9</sup> Sandblasting and etching with hydrofluoric acid are reliable methods to bond composite to porcelain.<sup>8-13</sup> However, studies on the repair strength of composite resins show considerable differences and contradictions in their results. Some previous studies have failed to show a positive effect of hydrofluoric acid, sandblasting or roughening with a bur on the repair strength of composite resin,<sup>7,14</sup> where most of them demonstrated a beneficial effect of such composite repair techniques.<sup>15-19</sup> The variation in results might be attributed to differences in the composition of the composite resins tested in these studies as repair strengths might be influenced by the type and amount of fillers present in the composite resin.

The most commonly used filler is silica which is bonded to composite resin matrix via silane therapy. There are, however, concerns regarding the possibility of the chemical bond hydrolysis between silica fillers and resin matrix;<sup>8,9</sup> therefore, zirconia-filled composite resins were introduced. Zirconia fillers are firmly retained in the resin matrix via surface micro-

porosities. Apparently, some surface treatments used for silica-filled composite resins might not be effective in zirconia-filled ones. For example, hydrofluoric acid and silane have no reaction with zirconia fillers.<sup>11,12</sup> Studies have revealed that functional organophosphate monomers can be used as primers with zirconia-filled porcelains<sup>13,20,21</sup> but no study has evaluated the efficacy of these primers in zirconia-filled composite resins. Functional monomers chemically bond to metal oxides on one side and copolymerize with resin monomers on the other.<sup>11</sup>

It is believed that despite similar resin matrix of the silica- and zirconia-filled composite resins, effect of surface treatments on repair bond strength would be different due to their dissimilar fillers.<sup>23</sup> Therefore, the aim of present study was to evaluate the effect of different surface treatments of alloy primer, silane and only surface roughening on repair micro-shear bond strength ( $\mu$ SBS) of silica- and zirconia-filled composite resins.

## Materials and Methods

Silica-filled composite, TPH Spectrum (Dentsply DeTrey GmbH, Konstanz, Germany), and zirconia-filled composite, Filtek Z250 (3M ESPE, St Paul, MN, USA), were used in the present study. For each composite resin, 27 blocks measuring  $15 \times 10 \times 1$  mm were made using an aluminum split mold. Composite resin blocks were cured step-by-step by Optilux 501 (Sybron Kerr, Danbury, CT, USA) at a light intensity of  $1000 \text{ mW/cm}^2$  for 20 seconds in a manner in which each curing area overlapped previously cured ones to some extent. Glass slabs were placed on the bottom as well as on the top of the mold prior to curing of composite layers to provide complete adaptability to the mold walls and produce a smooth and even surface. The light intensity was frequently monitored with Optilux Radiometer (Sybron Kerr, Danbury, CT, USA) to ensure adequate polymerization of all the specimens. Composite resin blocks were immersed in distilled water at  $37^\circ\text{C}$  for 24 hours and then thermocycled between  $5 \pm 5$  and  $55 \pm 5^\circ\text{C}$  for 5000 cycles<sup>23</sup> prior to surface roughening with a cylindrical medium-grit diamond bur (DIASWISS, Nyon, Switzerland). A new diamond bur was used for every 5 blocks. Then, the surfaces were etched using 35% phosphoric acid (Scotchbond Etchant, 3M ESPE, St Paul, MN, USA) for 20 seconds, thoroughly rinsed, and air-dried. Details of the materials used are presented in Table 1.

## Surface Treatments for Repair of Composite Resins

Twenty-seven blocks of each composite resin type

**Table 1. Study materials, their manufacturer and components**

Material	Manufacturer	Main Components
TPH Spectrum/Shade A1	Dentsply DeTrey, Konstanz, Germany	Bis-GMA, Bis-EMA, TEGDMA, barium aluminoborosilicate, silica
Filtek Z250/Shade A1	3M ESPE, St. Paul, MN, USA	Bis-GMA, UDMA, Bis-EMA Zirconia/silica Fillers (without silane treatment)
Scotchbond Etchant gel	3M ESPE, St. Paul, MN, USA	35% phosphoric acid, silicon dioxide
Margin Bond	Coltène/Whaledent AG, Switzerland	Bis-GMA, Bis-EMA, TEGDMA
Adper Single Bond Plus	3M ESPE, St. Paul, MN, USA	Bis-GMA, HEMA, polyalkenoiccopolymer, ethanol, photoinitiator
Clearfil SE Bond	Kuraray Medical Inc., Okayama, Japan	<i>SE Primer:</i> N,N-Diethanol-p-toluidine, MDP, HEMAhydrophilic dimethacrylate, DL-camphorquinone, water. <i>SE Bond:</i> N,N-Diethanol-p-toluidine, MDP, Bis-GMA, HEMA, hydrophobic dimethacrylate, DL camphorquinone, silanated,colloidal silica
Alloy Primer	Kuraray Medical Inc., Okayama, Japan	MDP
Porcelain Silane	Ultradent, Salt-Lake City, UT, USA	3-methacryloxy propyltrimethoxysilane, ethanol, water, acetic acid

**Bis-GMA:** bisphenol-glycidyl methacrylate, **Bis-EMA:** ethoxylatedbisphenol A glycol dimethacrylate, **TEGDMA:** triethylene glycol dimethacrylate, **UDMA:** urethane di-methacrylate, **HEMA:** 2-hydroxyethyl methacrylate **MDP:** 10-methacryloyloxydecyl dihydrogen phosphate.

were divided into 3 groups ( $n = 9$ ) according to their initial surface treatments:

1. Application of one layer of Alloy Primer (ALP) (Kuraray Medical Inc., Okayama, Japan) and air drying after 60 seconds
2. Application of one layer of coupling agent, Porcelain Silane (PS) (Ultradent, Salt-Lake City, UT, USA) and air drying after 60 seconds
3. No additional primer after surface roughening (SR)

#### *Application of Adhesive Systems*

Each group of surface-treated composite resin blocks was further subdivided into three subgroups ( $n = 3$ ) based on the adhesive system applied:

1. One layer of Margin Bond (MB) adhesive system (Coltène/Whaledent AG, Switzerland) and curing for 20 seconds
2. One layer of Clearfil SE Bond (CSE) adhesive system (Kuraray Medical Inc., Okayama, Japan) and curing for 20 seconds
3. Two layers of Single Bond (SB) adhesive system (3M ESPE, St Paul, MN, USA) and curing for 20 seconds, followed by 5 seconds of gentle air drying of each layer

#### *Preparation of Specimens for the $\mu$ SBS Test*

Iris cuts from silicone Tygon tubing (Small Parts Inc, Logansport, IN, USA) with an internal diameter of 1

mm and a height of approximately 0.5 mm were used to produce composite cylinders. On each composite resin block, four composite resin cylinders of the same type were placed with a distance of 2 mm from each other. The specimens were then thermocycled between  $5 \pm 5$  and  $55 \pm 5^\circ\text{C}$  for 1000 cycles. Then, the composite resin blocks were fixated on Universal Testing Machine (Hounsfield Test Equipment, Model H5-KS, Surray, UK) using cyanoacrylate glue (Mitreapel, Beta Chemical Ind. & Trade Inco. Co., Istanbul, Turkey); a micro-shear force with a cross-head speed of 0.5 mm/min was applied via copper wire placed around the composite cylinders.

#### *Statistical Analysis*

Since Kolmogorov-Smirnov test revealed normal distribution of data ( $p > 0.05$ ), multifactor ANOVA was used for statistical analysis. Furthermore, in cases in which the differences were statistically significant, after application of one-way ANOVA, a post-hoc Bonferroni test was used for two-by-two comparisons. Statistical significance was defined at  $p < 0.05$ .

#### *Results*

Table 2 presents the results of multifactor ANOVA which showed that the effect of composite resin type

**Table 2. Results of multi-factor ANOVA for mean of repair  $\mu$ SBS ( $p < 0.05$ )**

Source	Type III Sum of Squares	df	Mean Square	F	Sig.
Composite Type (CT)	4.134	1	4.134	0.238	0.627
Surface Treatment (ST)	150.612	2	75.306	4.328	0.015
Adhesive Systems (AS)	243.882	2	121.941	7.008	0.001
Interaction (CT-ST)	3.475	2	1.737	0.100	0.905
Interaction (CT-AS)	48.140	2	24.070	1.383	0.253
Interaction (ST-AS)	185.844	4	46.461	2.670	0.034
Interaction (CT-ST-AS)	310.114	4	77.528	4.456	0.002
Error	3288.607	189	17.400	—	—
Total	36153.446	207	—	—	—

was not significant ( $p > 0.05$ ), but the effects of the type of surface treatment ( $p = 0.01$ ) and the type of adhesive system ( $p = 0.01$ ) were significant on repair  $\mu$ SBS. In addition, the cumulative effects of composite resin type with the type of surface treatment ( $p > 0.05$ ) and the composite resin type with the type of the adhesive system ( $p > 0.05$ ) were not statistically significant. However, the cumulative effects of the adhesive system-surface treatment ( $p = 0.03$ ) and the composite resin type-adhesive system-surface treatments ( $p = 0.002$ ) were significant.

When repairing both composite resins with ALP or SP as surface treatments there were no statistically significant differences between different adhesive systems ( $p > 0.05$ ). However, with SR as surface treatment for both composite resins there were statistically significant differences between the adhesive systems ( $p_{TPH} < 0.001$ ,  $p_{Z250} = 0.03$ , one-way ANOVA). In this case, a Bonferroni test revealed the following: repair  $\mu$ SBS of TPH with SB = CSE ( $p > 0.05$ ), SB = CSE > MB ( $p < 0.001$ ) and repair  $\mu$ SBS of Z250 with SB > CSE ( $p = 0.02$ ), SB = MB and CSE = MB ( $p > 0.05$ ) (Table 3).

When repairing TPH using SB or CSE adhesive systems there were statistically significant differences between surface treatments applied ( $p_{SB} = 0.002$ ,  $p_{CSE} = 0.04$ , one-way ANOVA). However, for MB adhesive system there were statistically significant differences between surface treatments applied ( $p > 0.05$ ). Two-by-two comparison of surface treatments for SB adhesive system revealed the following: OR = PS ( $p > 0.05$ ), PS > ALP ( $p = 0.01$ ) and OR > ALP ( $p = 0.002$ ). Bonferroni test revealed the following: OR = PS ( $p > 0.05$ ), PS = ALP ( $p > 0.05$ ) and OR > ALP ( $p = 0.04$ ) with the use of CSE adhesive system. In case of repairing Z250 composite with all of the adhesive systems there were no statistically significant differences between different surface treatments applied ( $p > 0.05$ ) (Table 3).

## Discussion

Conventional shear and tensile tests have been criti-

cized for using relatively large bonded surfaces, over which stress distribution is likely to be uneven in relation to the density of intrinsic faults, possibly acting as stress raisers.<sup>24</sup> In this regard, the microtensile technique is considered more reliable, being able to more closely reflect the interfacial bond strength, as it offers more uniform stress distribution.<sup>25,26</sup> However, a high frequency of premature failures and large standard deviation values were reported.<sup>27</sup> In  $\mu$ SBS test, unlike microtensile technique, sectioning and trimming steps which may introduce early microcracking and pre-test failures within the specimen are avoided.<sup>28,29</sup> Therefore, it is more effective and reliable for evaluating bonding efficiency of adhesive systems.<sup>28</sup>

Clinically, aging is a result of exposure of composite materials to the oral environment, food and beverages of all kinds, and cyclic loading over a long period of service. This aging process will result in leaching of certain components out of the composite resin, water uptake in the resin matrix and along the resin-filler interface and wear of the surface due to loss of resin matrix and filler particles.<sup>30,31</sup> These changes can alter the composition of the material and will also affect the repair bond strength.<sup>32,33</sup> In a recent study, Ozcan et al<sup>15</sup> demonstrated that 'aging' using thermocycling for 5000 rounds was more efficacious than the effect of citric acid (3.5% for 1 week) and storage in boiling water for 8 hours. Therefore, in the present study the technique of thermocycling for 5000 rounds was chosen.

The present study was conducted to determine the effect of different combinations of surface treatments and adhesive systems on repair  $\mu$ SBS of two types of composite resins with dissimilar filler types: Filtek Z250, zirconia-filled and Spectrum TPH, silica-filled.

After aging in a humid environment, the water saturation of the composite resin was accomplished, and the monomer functional group's radical activity diminished.<sup>33</sup> Therefore, the surfaces of the aged composite resins need to be refreshed somehow. The

**Table 3. Means and standard deviations (Mean  $\pm$  SD) of the repair  $\mu$ SBS values (MPa)**

Composite Resin	Surface Treatments	Adhesive Systems		
		Single Bond	Clearfil SE Bond	Margin Bond
TPH Spectrum	Alloy Primer	aA*9.6 $\pm$ 3.8	4.4 aA11.5 $\pm$	Aa12.8 $\pm$ 3.9
	Porcelain Silane	bA14.0 $\pm$ 3.2	14.3 $\pm$ 4.0 abA	4.3 aA $\pm$ 10.3
	Surface Roughening	bA15.0 $\pm$ 3.6	bA15.7 $\pm$ 3.09	2.7 aB $\pm$ 9.2
	Alloy Primer	3.1 aAB*11.1 $\pm$	5.5 aA13.1 $\pm$	aB $\pm$ 3.28.7
Filtek Z250	Porcelain Silane	4.0 abA13.7 $\pm$	4.8 aA $\pm$ 13.4	aA $\pm$ 4.411.6
	Surface Roughening	6.0 bA15.8 $\pm$	3.7 aB $\pm$ 10.2	5.0 aAB $\pm$ 12.2

Same lower letters indicate no significant difference in bond strength between different surface treatment groups when using similar adhesives in each composite type.

Same capitals indicate no significant difference of bond strength between different adhesive groups when using similar surface treatments in each composite type.

use of an intermediate low-viscosity resin can be considered a necessary step in composite resin repair, to enhance the bond by promoting chemical coupling to the resin matrix, bonding to the exposed fillers, or micromechanical retention through monomer penetration into the matrix microcracks.<sup>15,34</sup>

To this end, three adhesive systems with different characteristics were used in this study: Margin Bond, an enamel bonding containing hydrophobic monomers; Single Bond, an etch-and-rinse adhesive system, containing hydrophobic monomers together with hydrophilic monomers; and Clearfil SE Bond, a two-bottle (primer + bond) self-etch adhesive system, containing functional organophosphate monomers.

On repair protocol, in all the groups surface roughening was performed with a diamond bur followed by etching with phosphoric acid prior to the application of surface primers. According to Loomans et al, besides the superficial cleaning capacity of phosphoric acid, which removes debris and grinding debris from the resin composite surface,<sup>22,35</sup> phosphoric acid alone had no effect on surface roughness.<sup>22</sup> Moreover, etching with phosphoric acid might also activate the reactivity between silica or zirconia surface and PS or ALP which were used as surface primers in this study.<sup>35</sup> Furthermore, it should be pointed out that in the present study, for this reason and also because of strong hydrophilic nature of primer component of CSE adhesive system, which might be incompatible with mainly hydrophobic surface of composite resin, primer component of this adhesive was not applied prior to the application of bonding. We were also inspired to use MB because of mainly hydrophobic nature of composite resin surface.

In the present study, repair  $\mu$ SBS of TPH using SB adhesive system was higher when OR or PS were applied as surface treatment compared to that of ALP. This means that the aforementioned combination using PS as chemical primer after surface roughening does not cause any improvement in repair  $\mu$ SBS and use of ALP may decrease it. ALP is a metal primer containing MDP without any methyl-methacrylate-based ingredient to bond with SB. When repairing TPH using CSE adhesive system repair  $\mu$ SBS with OR as surface treatment was similar to PS and PS was similar to ALP. However, repair  $\mu$ SBS with OR was significantly higher than that with ALP. These mean that in such a combination there is no need for applying chemical primers after surface roughening in order to increase repair  $\mu$ SBS. However, in contrast to SB, repair  $\mu$ SBS with PS

was similar to ALP. This might be attributed to compatibility and probability of its chemical bonding to ALP primer due to the presence of MDP in their composition. Furthermore, when repairing TPH with MB adhesive system there were no significant differences between surface treatments, either.

Another finding of the present study was that no differences were found when using different combinations of adhesive systems and surface treatments in order to repair Z250 composite resin. The reason may be that since fillers comprise only a small percentage of the composite surface and the greater proportion is devoted to resin, surface treatments do not appear as efficient as in the case of porcelains.<sup>22</sup> That is why adhesives such as MB, which is only composed of hydrophobic monomers without any functional groups, could wet composite surface properly.

Previously, several methods have been used in order to roughen aged composite resin surfaces. Hydrofluoric acid is corrosive and a contact poison; a meticulous application technique is needed to prevent detrimental side effects, such as acid burns and necrosis of the underlying soft tissues.<sup>36-38</sup> Therefore, this material is less safe for intra-oral application. On the other hand, a major disadvantage of sandblasting is the aerosol of fine abrasive particles that will contaminate a wide area of the operatory, which might be harmful for patients and operators.<sup>39</sup>

However, roughening with diamond bur, used in this study, might be the simplest, the most feasible and aggressive of them. Contradictory results have been reported with the use of diamond burs for preparing composite surfaces prior to bonding.<sup>7,33,40,41</sup> Nevertheless, based on the results of this study and similar to those of a study by Soderholm et al,<sup>29</sup> surface roughening with diamond bur followed by acid etching with phosphoric acid and applying an adhesive system with good wetting properties seems to be the most feasible technique for dentists to use. Furthermore, applying primers such as PS or ALP, which contain functional monomers in order to react with filler particles, does not have any significant effects on the repair  $\mu$ SBS, regardless of the filler type.

However, it should be noted that the bond strength values do not necessarily correlate with microleakage, at least in the laboratory.<sup>42</sup> Therefore, it is recommend that microleakage of repaired interfaces and SEM observations of debonded surfaces be evaluated in future. Furthermore, no data appears to exist on the correlation between bond strength of repaired composite resins and clinical performance.



## Conclusion

Although repair  $\mu$ SBS values of both silica- and zirconia-filled composite resins were similar, using different combinations of surface treatments and/or adhesive systems affected their repair  $\mu$ SBS values differently. However, based on this study and within the limitations described, surface roughening with diamond bur followed by acid etching with phosphoric acid and applying an adhesive system with good wetting properties is recommended for repairing aged composite resins, regardless of the filler type.

## References

1. Tyas MJ, Anusavice KJ, Frencken JE, Mount GJ. Minimal intervention dentistry-a review. FDI Commission Project 1-97. *Int Dent J* 2000;50:1-12.
2. Li J. Effects of surface properties on bond strength between layers of newly cured dental composites. *J Oral Rehabil* 1997; 24:358-60.
3. Suh BI. Oxygen-inhibited layer in adhesion dentistry. *J Esthet Restor Dent* 2004;16:316-23.
4. Sehgal A, Rao YM, Joshua M, Narayanan LL. Evaluation of the effects of the oxygen-inhibited layer on shear bond strength of two resin composites. *J Conserv Dent* 2008;11:159-61.
5. Kao EC, Pryor HG, Johnston WM. Strength of composites repaired by laminating with dissimilar composites. *J Prosthet Dent* 1988;60:328-33.
6. Harry FA. Tooth-Colored Restoratives: *Principles and Techniques*, 9<sup>th</sup> ed. Canada: Quintessence;2002.
7. Brosh T, Pilo R, Bichacho N, Blutstein R. Effect of combinations of surface treatments and bonding agents on the bond strength of repaired composites. *J Prosthet Dent* 1997;77:122-6.
8. Takeshige F, Kawakami Y, Hayashi M, Ebisu S. Fatigue behavior of resin composites in aqueous environments. *Dent Mater* 2007;23:893-9.
9. Ferracane JL, Hopkin JK, Condon JR. Properties of heat-treated composites after aging in water. *Dent Mater* 1995;11:354-8.
10. Marshall SJ, Bayne SC, Baier R, Tomsia AP, Marshall GW. A review of adhesion science. *Dent Mater* 2010; 26:11-6.
11. Kern M, Wegner SM. Bonding to zirconia ceramic: adhesion methods and their durability. *Dent Mater* 1998; 14: 64-71.
12. Dérand P, Dérand T. Bond strength of luting cements to zirconium oxide ceramics. *Int J Prosthodont* 2000; 13:131-5.
13. Aboushelib MN, Matinlinna JP, Salameh Z, Ounsi H. Innovations in bonding to zirconia-based materials: Part I. *Dent Mater* 2008; 24:1268-72.
14. Lucena-Martín C, González-Lo'pez S, Navajas-Rodríguez de Mondelo JM. The effect of various surface treatments and bonding agents on the repaired strength of heat-treated composites. *J Prosthet Dent* 2001;86:481-8.
15. Özcan M, Barbosa SH, Melo RM, Galhano GAP, Bottino MA. Effect of surface conditioning methods on the microtensile bond strength of resin composite to composite after aging conditions. *Dent Mater* 2007;23:1276-82.
16. Passos SP, Özcan M, Vanderlei AD, Leite FP, Kimpara ET, Bottino MA. Bond strength durability of direct and indirect composite systems following surface conditioning for repair. *J Adhes Dent* 2007;9:443-7.
17. Rodrigues Jr SA, Ferracane JL, Della Bona A. Influence of surface treatments on the bond strength of repaired resin composite restorative materials. *Dent Mater* 2009;25:442-51.
18. Rinastiti M, Özcan M, Siswomihardjo W, Busscher HJ. Immediate repair bond strengths of microhybrid, nanohybrid and nanofilled composites after different surface treatments. *J Dent* 2010;38:29-38.
19. Hannig C, Laubach S, Hahn P, Attin T. Shear bond strength of repaired adhesive filling materials using different repair procedures. *J Adhes Dent* 2006;8:35-40.
20. Aboushelib MN, Mirmohamadi H, Matinlinna JP, Kukk E, Ounsi HF, Salameh Z. Innovations in bonding to zirconia-based materials. Part II: Focusing on chemical interactions. *Dent Mater* 2009;25:989-93.
21. Mirmohammadi H, Aboushelib MN, Salameh Z, Feilzer AJ, Kleverlaan CJ. Innovations in bonding to zirconia based ceramics: Part III. Phosphate monomer resin cements. *Dent Mater* 2010;26:786-92.
22. Loomans BAC, Cardoso MV, Opdam NJM, Roeters FJM, De Munck J, Huysmans MCDNJM, Van Meerbeek B. Surface roughness of etched composite resin in light of composite repair. *J Dent* 2011;39:499-505.
23. Antonucci JM, Dickens SH, Fowier BO, Xu HHK, McDonough WG. Chemistry of silanes: interfaces in dental polymers and composites. *J Res Natl Inst Stand Technol* 2005;110:541-58.
24. Sano H, Shono T, Sonoda H, Takatsu T, Ciucchi B, Carvalho R, Pashley DH. Relationship between surface area for adhesion and tensile bond strength--evaluation of a microtensile bond test. *Dent Mater* 1994;10:236-40.
25. Perdigão J, Gomes G, Gondo R, Fundingsland JW. In vitro bonding performance of all-in-one adhesives. Part I-microtensile bond strengths. *J Adhes Dent* 2006;8:367-73.
26. Ferrari M, Goracci C, Sadek F, Eduardo P, Cardoso C. Microtensile bond strength tests: scanning electron microscopy evaluation of sample integrity before testing. *Eur J Oral Sci* 2002;110:385-91.
27. Goracci C, Sadek FT, Monticelli F, Cardoso PE, Ferrari M. Influence of substrate, shape, and thickness on microtensile specimens' structural integrity and their measured bond strengths. *Dent Mater* 2004; 20:643-54.
28. Ishikawa A, Shimada Y, Foxton RM, Tagami J. Microtensile and micro-shear bond strengths of current self-etch adhesives to enamel and dentin. *Am J Dent* 2007;20:161-6.
29. Söderholm KJ, Roberts MJ. Variables influencing the repair strength of dental composites. *Scand J Dent Res* 1991; 99:173-80.
30. Tarumi H, Torii M, Tsuchitani Y. Relationship between particle size of barium glass filler and water sorption of light-cured composite resin. *Dent Mater* 1995;14:37-44.
31. Finer Y, Santerre JP. Salivary esterase activity and its association with the biodegradation of dental composites. *J Dent Res* 2004;83:22-6.
32. Brendeke J, Özcan M. Effect of physicochemical aging conditions of the composite-composite repair bond strength. *J Adhes Dent* 2007;9:399-406.
33. Bonstein T, Garlapo D, Donarummo J Jr, Bush PJ. Evaluation of varied repair protocols applied to aged composite resin. *J Adhes Dent* 2005;7:41-49.
34. Kupiec KA, Barkmeier WW. Laboratory evaluation of surface treatments for composite repair. *Oper Dent* 1996;21:59-62.
35. Loomans BA, Cardoso MV, Roeters FJ, Opdam NJ, De Munck J, Huysmans MC, Van Meerbeek B. Is there one optimal repair technique for all composites? *Dent Mater*

- 2011;27:701-9.
36. Sheridan RL, Ryan CM, Quinby WC, Blair J, Tompkins RG, Burke JF. Emergency management of major hydrofluoric acid exposures. *Burns* 1995;21:62-4.
37. Gallerani M, Bettoli V, Person L, Manfredini R. Systemic and topical effects of intradermal hydrofluoric acid. *Am J Emerg Med* 1989;16:521-2.
38. Asvesti C, Guadagni F, Anastasiadis G, Zakopoulou N. Hydrofluoric acid burns. *Cutis* 1997;59:306-8.
39. Roeters FJM. A simple method to protect patient and environment when using sandblasting for intraoral repair. *J Adhes Dent* 2000;2:235-8.
40. Shahdad SA, Kennedy JG. Bond strength of repaired anterior composite resins: an in vitro study. *J Dent* 1998; 26:685-94.
41. Papacchini F, Dall'Oca S, Chieffi N, Goracci C, Sadek FT, Suh BI, et al. Composite-to-composite microtensile bond strength in the repair of a microfilled hybrid resin: effect of surface treatment and oxygen inhibition. *J Adhes Dent* 2007;9:25-31.
42. Tagami J, Nikaido T, Nakajima M, Shimada Y. Relationship between bond strength tests and other in vitro phenomena. *Dent Mater* 2010;26:e94-e99.